

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

	<h3 style="margin: 0;">Haven Automation Limited</h3>	
	<p>Issue No: 022 Issue date: 11 August 2004</p>	
	<p>Measurement House Kingsway Fforestfach Swansea SA5 4EX</p>	<p>Contact: Mr J P Vaughan Tel: +44 (0)1792 588722 Fax: +44 (0)1792 582624 E-Mail: mail@haven.co.uk Website: www.haven.co.uk</p>

SUMMARY OF ACCREDITATION

Calibration performed on permanent laboratory premises

ELECTRICAL

Ammeters, ac
Ammeters, dc
Bridges and similar instruments, resistance
Calibrators, multimeter
Capacitance meters
Frequency counters
Frequency meters
Insulation testers
Multimeters, analogue
Multimeters, digital
Potentiometers
Power supply units
Resistance boxes
Resistance meters
Resistors, dc
Shunts
Voltmeters, ac
Voltmeters, dc

PRESSURE

Gas gauge, indicating instruments
Gas gauge, devices with an electrical output
Gas negative gauge, indicating instruments
Gas negative gauge, devices with an electrical output
Hydraulic gauge, indicating instruments
Hydraulic gauge, devices with an electrical output

TEMPERATURE

Block 'calibrators'
Liquid baths
Resistance thermometers
Temperature indicators and recorders, with temperature sensor(s)
Temperature indicators and recorders, electrical calibration without sensor
Thermocouples, base and noble metal types, e.g. K, N, T, R, S



Calibration performed on permanent laboratory premises

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Haven Automation Limited

Issue No: 022 Issue date: 11 August 2004

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks
ELECTRICAL			
DC RESISTANCE			
Measurement	Up to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 k Ω 2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 2 M Ω 200 k Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 200 M Ω 200 M Ω to 1 G Ω	30 ppm + 100 $\mu\Omega$ 50 ppm 30 ppm 30 ppm 30 ppm 50 ppm 50 ppm 100 ppm 350 ppm 0.65%	
Generation	Up to 40 Ω 40 Ω to 400 Ω 400 Ω to 4 k Ω 4 k Ω to 40 k Ω 40 k Ω to 400 k Ω 400 k Ω to 4 M Ω 4 M Ω to 40 M Ω 40 M Ω to 400 M Ω	250 ppm + 100 $\mu\Omega$ 100 ppm 100 ppm 100 ppm 100 ppm 120 ppm 300 ppm 300 ppm	
DC VOLTAGE			
Measurement	Up to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1 kV	3 μ V 15 ppm 15 ppm 40 ppm 40 ppm	
Generation	Up to 320 mV 320 mV to 3.2 V 3.2 V to 32 V 32 V to 320 V 320 V to 1050 V	50 ppm + 2 μ V 140 ppm 140 ppm 140 ppm 140 ppm	
DC CURRENT			
Measurement	Up to 20 mA 20 mA to 200 mA 200 mA to 2 A	70 ppm + 10 nA 120 ppm 220 ppm	
Generation	Up to 320 μ A 320 μ A to 3.2 mA 3.2 mA to 32 mA 32 mA to 320 mA 320 mA to 3 A 3A to 10 A	120 ppm + 6 nA 300 ppm 300 ppm 0.065% 0.07% 0.065%	
AC VOLTAGE			



Calibration performed on permanent laboratory premises

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Haven Automation Limited

Issue No: 022 Issue date: 11 August 2004

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks
Generation	320 mV to 300 V 40 Hz to 30 kHz	0.05%	
	300 V to 750 V 40 Hz to 10 kHz	0.06%	
	750 V to 1050 V 40 Hz 60 10 kHz	0.1%	
AC CURRENT			
Generation	32 mA to 320 mA 10 Hz to 110 Hz 110 Hz to 3 kHz	0.04% 0.13%	
	320 mA to 3 A 40 Hz to 110 Hz 110 Hz to 3 kHz	0.05% 0.08%	
	3 A to 10 A 40 Hz to 110 Hz 110 Hz to 3 kHz	0.08% 0.18%	
CAPACITANCE			
Generation	1 nF to 4 nF 4 nF to 40 nF 40 nF to 400 nF 400 nF to 4 µF 4 µF to 40 µF 40 µF to 400 µF 400 µF to 4 mF 4 mF to 30 mF	0.3% 0.2% 0.2% 0.2% 0.2% 0.2% 0.35% 0.35%	
FREQUENCY			
Generation	0.5 Hz to 200 kHz	0.002% + 0.01 Hz	
PRESSURE			NOTES
<u>Hydraulic pressure (gauge)</u>			1 Calibrations may be undertaken expressed in other units of pressure as required.
Calibration of pressure indicating instruments and gauges	550 kPa to 6 MPa 6 MPa to 60 MPa	0.02% 0.015%	
<u>Gas pressure (gauge)</u>			2 Calibration of pressure measuring devices with an electrical output may be undertaken.
Calibration of pressure indicating instruments and gauges	-90 kPa to -5 kPa 5 kPa to 35 kPa 35 kPa to 42 kPa 42 kPa to 700 kPa 700 kPa to 3.45 MPa	0.025% 0.02% 0.01% 0.012% 0.02%	
TEMPERATURE			



Calibration performed on permanent laboratory premises

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Haven Automation Limited

Issue No: 022 Issue date: 11 August 2004

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks
Resistance Thermometers	-70 to below 0 °C 0 °C	0.07 °C 0.06 °C	
	above 0 to 100 °C above 100 to 300 °C above 300 to 600 °C	0.10 °C 0.15 °C 0.22 °C	
Thermocouples (Base and noble metals)	below 0 °C down to -70 °C 0 °C up to +50 °C above 50 °C up to 300 °C above 300 °C up to 600 °C	0.5 °C 0.3 °C 0.5 °C 1.0 °C	
Electronic thermometers with sensors		As for sensor plus: ½ scale division	
Analogue	As for Sensor		
Digital	As for Sensor	1 lsd	
Solid state temperature sensors with indicators	-70 °C up to 150 °C	0.3 °C + 1 lsd	
Temperature Indicators (electrical simulation test)	-50 to 600 °C above 600 °C up to 1768 °C	0.6 °C 1.0 °C	
Metal Block calibrators	-50 to 300 °C above 300 °C up to 600 °C	0.3 °C 1.0 °C	
Liquid Baths	0 °C to 250 °C	0.2 °C	
END			